

CERTIFICATE

SAMPLER | VIALS | TEFLON WARE
HYDROTHERMAL | SS CLEAN ROOM ITEMS



MATERIAL SPECIFICATION CERTIFICATE

Part Number : 24131

MSC No. : SAMPLE

Part Name : Bubble Trap



Dia : 100 mm

Flow Rate : 4 litres / minute

Tube : Borosilicate 3.3 glass (Acrylic Optional)

Operating Pressure : 7.0 bar / 102 psi

Base Inlet / Outlet Connectors : Tri Clamp Ferrules (1/2" / 3/4" / 1.0" / 1.1/2" / 2.0")

Lid Controls : Tri Clamp Ferrules 2 ports / 3 ports (1/2" / 3/4" / 1.0" / 1.1/2" / 2.0")

Material : Stainless Steel 316 / 316 L



JN SCIENCETECH

Works : # 7-4-232, Geetha Nagar, Ferozguda,
Balanagar, Hyderabad - 500 011.

✉ info@jnsciencetech.com | info@jnsciencetech.in

🌐 www.jnsciencetech.com | www.jnsciencetech.in

ISO 9001:2015 CERTIFIED